

United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.	
10/608,257	06/27/2003	Guy P. Vanney	11530US.01	. 8338	
33486 75	590 11/30/2006		EXAMINER		
HEIMBECHER & ASSOC., LLC			ROLLINS, ROSILAND STACIE		
P O BOX 33 HAMEL, MN	55340-0033		ART UNIT	PAPER NUMBER	
,			3739		
			DATE MAILED: 11/30/2006		

Please find below and/or attached an Office communication concerning this application or proceeding.

• •					X				
Office Action Summary The MAILING DATE of this communication ap		Application	No.	Applicant(s)					
		10/608,257		VANNEY ET AL.					
		Examiner		Art Unit					
		Rosiland S.		3739					
Period for Reply	ATE OF this communication app	pears on the c	over sneet with the co	orrespondence ad	aress				
WHICHEVER IS LON - Extensions of time may be a after SIX (6) MONTHS from - If NO period for reply is spec - Failure to reply within the se	GER, FROM THE MAILING DAVAILABLE UNDER THE MAILING THE MAI	ATE OF THIS 36(a). In no event, will apply and will e c, cause the applica	COMMUNICATION however, may a reply be tim xpire SIX (6) MONTHS from to tion to become ABANDONED	I. lety filed the mailing date of this co (35 U.S.C. § 133)					
Status									
1) Responsive to d	ommunication(s) filed on 15 Au	ugust 2006.							
2a)⊠ This action is FI	☐ This action is FINAL . 2b)☐ This action is non-final.								
	cation is in condition for allowar		·		e merits is				
closed in accord	closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.								
Disposition of Claims									
4a) Of the above 5) ⊠ Claim(s) <u>1,5-9,1</u> 6) ⊠ Claim(s) <u>19</u> is/a 7) □ Claim(s)	/are pending in the application. e claim(s) 3,4,10 and 13 is/are v 1,12 and 14 is/are allowed. re rejected. is/are objected to. are subject to restriction and/or	withdrawn fro							
Application Papers									
10) The drawing(s) f Applicant may no Replacement draw	is objected to by the Examine iled on is/are: a) accept request that any objection to the owing sheet(s) including the correction is objected to by the Examine.	epted or b) drawing(s) be tion is required	held in abeyance. See if the drawing(s) is obj	e 37 CFR 1.85(a). ected to. See 37 CF					
Priority under 35 U.S.C.	§ 119								
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 									
	Patent Drawing Review (PTO-948) externent(s) (PTO-1449 or PTO/SB/08)) Interview Summary (Paper No(s)/Mail Da) Notice of Informal Pa) Other:	te	O-152)				

DETAILED ACTION

Claim Rejections - 35 USC § 102

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claim 19 is rejected under 35 U.S.C. 102(b) as being anticipated by Bednarek et al. (US 6120500). Bednarek et al. dislcose an ablation catheter comprising a catheter shaft, the shaft comprising a proximal portion, a distal portion, the distal portion comprising an outer peripheral wall having an active region, wherein the distal portion has a cross-sectional configuration along the active region and the cross-sectional configuration is adapted to bias the active region against the tissue to be ablated.

Allowable Subject Matter

Claims 1, 5-9 and 11, 12, 14-18 are allowed.

Response to Arguments

Applicant's arguments filed 8/15/06 have been fully considered but they are not persuasive. Applicant argues that Bednarek does not disclose a cross-sectional configuration having an outer periphery that is adapted to bias the outer peripheral wall of the catheter against the tissue to be ablated. Figure 1 of Bednarek et al. clearly illustrates that the configuration of the catheter, which includes its cross section, is inherently adapted to bias the outer peripheral wall against the tissue to be ablated.

Conclusion

Art Unit: 3739

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Rosiland S. Rollins whose telephone number is (571) 272-4772. The examiner can normally be reached on Mon.-Fri. 9:00 AM - 6:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C. Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/608,257

Art Unit: 3739

Page 4

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Rosiland S Rollins
Primary Examiner
Art Unit 3739